

**Europe / Middle-East / Africa**

Microlife AG  
Espanstrasse 139  
9443 Widnau / Switzerland  
Tel. +41 / 71 727 70 30  
Fax +41 / 71 727 70 39  
Email admin@microlife.ch  
www.microlife.com

**Asia**

Microlife Corporation  
9F, 431, RuiGuang Road, NeiHu  
Taipei 11492, Taiwan, R.O.C.  
Tel. +886-2-87971288  
Fax +886-2-87971283  
Email: service@microlife.com.tw  
www.microlife.com

CE0044

IB BP B6 Advanced Connect EN-IN 0920

**microlife**<sup>®</sup>

**B6** Advanced Connect




**Microlife BP B6 Advanced Connect**


EN → 1


Preparation

- 

90° Sit on a back-supported chair and keep your legs uncrossed.
- 


Avoid thick or close-fitting garments on the upper arm.
- 

Place the artery-mark on the cuff over your artery.
- 

Fit the cuff closely, but not too tight.
- 

Position the cuff 2-3 cm above your elbow.
- 

Keep your arm still and do not speak during the measurement.



1.



Avoid eating, bathing, smoking or caffeine (approx. 30 min).

2.



Relax 5-10 Mins

Avoid activity and relax for 5-10 min.

3.



Measure before medication intake.

Name of Purchaser

---

Serial Number

---

Date of Purchase

---

Specialist Dealer

---

- ① ON/OFF button
- ② Display
- ③ M-button (memory)
- ④ Time Button
- ⑤ User Button
- ⑥ Cuff Socket
- ⑦ MAM Switch
- ⑧ Traffic Light Indicator
- ⑨ USB Port
- ⑩ Mains Adapter Socket
- ⑪ Battery Compartment
- ⑫ Cuff
- ⑬ Cuff Tube
- ⑭ Cuff Connector

## Display

- ⑮ Active Bluetooth®
- ⑯ Systolic Value
- ⑰ Diastolic Value
- ⑱ Pulse Rate
- ⑲ Battery Display
- ⑳ Cuff Fit Check
  - A: Suboptimal Cuff Fit
  - B: Arm Movement Indicator «**Err 2**»
  - C: Cuff Pressure Check «**Err 3**»
- ㉑ Cuff Signal Indicator «**Err 1**»
- ㉒ Pulse Arrhythmia Indicator (PAD)
- ㉓ Atrial Fibrillation Indicator (AFIB)
- ㉔ User Indicator
- ㉕ MAM Mode
- ㉖ Stored Value
- ㉗ Clinical Blood Pressure Average «MyBP»
- ㉘ Date/Time
- ㉙ Pulse Indicator
- ㉚ Average Indicator «MyCheck»



Follow Instructions for Use. This document provides important product operation and safety information regarding this Blood Pressure Monitor. Please read this document thoroughly before using the device and keep for future reference.



Type BF applied part



Keep dry

## Intended use:

This oscillometric blood pressure monitor is intended for measuring non-invasive blood pressure in people aged 12 years or older. It is clinically validated in patients with hypertension, hypotension, diabetes, pregnancy, pre-eclampsia, atherosclerosis, end-stage renal disease, obesity and the elderly.

The device can detect an irregular pulse suggestive of Atrial Fibrillation (AF). Please note that the device is not intended to diagnose AF. A diagnosis of AF can only be confirmed by ECG. The patient is advised to see a physician.

## Dear Customer,

This device was developed in collaboration with physicians and clinical tests carried out prove its measurement accuracy to be of a very high standard.\*

Microlife AFIBsens™ is the world's leading digital blood pressure measurement technology for the detection of atrial fibrillation (AF) and arterial hypertension. These are the two top risk factors of getting a stroke or heart disease. It is important to detect AF and hypertension at an early stage, even though you may not experience any symptoms. AF screening in general and thus also with the Microlife AFIBsens™ algorithm, is recommended for people of 65 years and older. The AFIBsens™ algorithm indicates that atrial fibrillation may be present. For this reason, it is recommended that you visit your doctor when the device gives an AFIB signal during your blood pressure measurement. The AFIBsens™ algorithm of Microlife has been clinically investigated by several prominent clinical investigators and showed that the device detects patients with AF at a certainty of 97-100%.<sup>1,2</sup>

If you have any questions, problems or want to order spare parts please contact your local Microlife-Customer Service. Your dealer or pharmacy will be able to give you the address of the Microlife

dealer in your country. Alternatively, visit the internet at [www.microlife.com](http://www.microlife.com) where you will find a wealth of invaluable information on our products.

Stay healthy – Microlife !

*\* This device uses the same measuring technology as the award winning «BP 3BTO-A» model tested according to the British Hypertension Society (BHS) protocol.*

<sup>1</sup> Kearley K, Selwood M, Van den Bruel A, Thompson M, Mant D, Hobbs FR et al.: Triage tests for identifying atrial fibrillation in primary care: a diagnostic accuracy study comparing single-lead ECG and modified BP monitors. *BMJ Open* 2014; 4:e004565.

<sup>2</sup> Wiesel J, Arbesfeld B, Schechter D: Comparison of the Microlife blood pressure monitor with the Omron blood pressure monitor for detecting atrial fibrillation. *Am J Cardiol* 2014; 114:1046-1048.

## Table of Contents

### 1. Appearance of the Atrial Fibrillation Indicator for early Detection

- What is Atrial Fibrillation (AF)?
- Who should be screened for Atrial Fibrillation?
- Risk factors you can control

### 2. Using the Device for the First Time

- Inserting the batteries
- Setting the date and time
- Selecting the correct cuff
- Selecting the user
- Selecting standard or MAM mode

### 3. Checklist for Taking a Reliable Measurement

#### 4. Taking a Blood Pressure Measurement

- Manual inflation
- How not to store a reading
- How do I evaluate my blood pressure?
- Average Indicator «MyCheck»
- Appearance of the Pulse Arrhythmia (PAD)

#### 5. Data Memory

- Viewing the average of the last 28 days
- Viewing the clinical blood pressure average «MyBP»
- Viewing the stored single values
- Clearing all values

#### 6. Battery Indicator and Battery change

- Low battery
- Flat battery – replacement
- Which batteries and which procedure?
- Using rechargeable batteries

### 7. Using a Mains Adapter

### 8. Bluetooth® Function

### 9. PC-Link Functions

### 10. Error Messages

### 11. Safety, Care, Accuracy Test and Disposal

- Safety and protection
- Device care
- Cleaning the cuff
- Accuracy test
- Disposal

### 12. Guarantee

### 13. Technical Specifications

## 1. Appearance of the Atrial Fibrillation Indicator for early Detection

This device is able to detect atrial fibrillation (AF) during blood pressure measurement in both standard mode and MAM mode. This symbol  indicates that atrial fibrillation was detected during the measurement. Please refer to the next paragraph for information regarding the consultation with your doctor.

### Information for the doctor on frequent appearance of the atrial fibrillation indicator

This device is an oscillometric blood pressure monitor that also analyses pulse irregularity during measurement. The device is clinically tested.

The AFIB symbol is displayed after the measurement, if atrial fibrillation occurred during measuring. If the AFIB symbol appears after having performed a full blood pressure measurement, the patient is advised to perform another measurement. If the AFIB symbol appears again, we recommend the patient to seek medical advice.

If the AFIB-symbol appears on the screen of the blood pressure monitor, it indicates the possible presence of atrial fibrillation. The atrial fibrillation diagnosis however, **must** be made by a **cardiologist** based on ECG interpretation.

-  Keep the arm still during measuring to avoid false readings.
-  This device may not or wrongly detect atrial fibrillation in people with pacemakers or defibrillators.
-  In the presence of atrial fibrillation the diastolic blood pressure value may not be accurate.
-  In the presence of atrial fibrillation using MAM-mode is recommended for more reliable blood pressure measurement.

### What is Atrial Fibrillation (AF)?

Normally, your heart contracts and relaxes to a regular beat. Certain cells in your heart produce electrical signals that cause the heart to contract and pump blood. Atrial fibrillation occurs when rapid, disorganized electrical signals are present in the heart's two upper chambers, called the atria; causing them to contract irregularly (this is called fibrillation). Atrial fibrillation is the most common form of heart arrhythmia. It often causes no symptoms, yet it significantly increases your risk of stroke. You'll need a doctor to help you control the problem.

### Who should be screened for Atrial Fibrillation?

AF screening is recommended for people over 65 years of age, since the chance of having a stroke increases with age. AF screening is also recommended for people from the age of 50 years who have high blood pressure (e.g. SYS higher than 159 or DIA higher than 99) as well as those with diabetes, coronary heart failure or for those who have previously had a stroke.

In young people or in pregnancy AF screening is not recommended as it could generate false results and unnecessary anxiety. In addition, young individuals with AF have a low risk of getting stroke as compared to elder people.

-  Microlife AFIBsens technology provides a convenient way to screen for AF  
Knowing your blood pressure and knowing whether you or your family members have AF can help reduce the risk of stroke.  
Microlife AFIBsens technology provides a convenient way to screen for AF whilst taking blood pressure measurement.

### Risk factors you can control

Early diagnosis of AF followed by adequate treatment can significantly reduce the risk of getting stroke. Knowing your blood pressure and knowing whether you have AF is the first step in proactive stroke prevention.

For more information visit our website: [www.microlife.com/afib](http://www.microlife.com/afib).

## 2. Using the Device for the First Time

### Inserting the batteries

BP B6 Advanced Connect

After you have unpacked your device, first insert the batteries. The battery compartment (1) is on the bottom of the device. Insert the batteries (4 x 1.5 V, size AA), thereby observing the indicated polarity.

### Setting the date and time

1. After the new batteries are fitted, the year number flashes in the display. You can set the year by pressing the M-button (3). To confirm and then set the month, press the time button (4).
2. Press the M-button to set the month. Press the time button to confirm and then set the day.
3. Follow the instructions above to set the day, hour and minutes.
4. Once you have set the minutes and pressed the time button, the date and time are set and the time is displayed.
5. If you want to change the date and time, press and hold the time button down for approx. 3 seconds until the year number starts to flash. Now you can enter the new values as described above.

### Selecting the correct cuff

Microlife offers different cuff sizes. Select the cuff size to match the circumference of your upper arms (measured by close fitting in the centre of the upper arm).

| Cuff size | for circumference of upper arm |
|-----------|--------------------------------|
| S         | 17 - 22 cm                     |
| M         | 22 - 32 cm                     |
| M - L     | 22 - 42 cm                     |
| L         | 32 - 42 cm                     |
| L - XL    | 32 - 52 cm                     |

 Only use Microlife cuffs.

- ▶ Contact your local Microlife Service if the enclosed cuff (12) does not fit.
- ▶ Connect the cuff to the device by inserting the cuff connector (14) into the cuff socket (6) as far as it will go.

### Selecting the user

This device allows to store the results for 2 individual users.

- ▶ Select the intended user (user 1 or user 2 (24)) by pressing the user button (5).

 Before each measurement, ensure that the correct user is selected.

### Selecting standard or MAM mode

Before each measurement, select standard (single measurement) or MAM mode (automatic double measurement). In MAM mode, 2 measurements are automatically taken in succession and the result is then automatically analysed and displayed. Because the blood pressure constantly fluctuates, a result obtained in this way is more reliable than when a single measurement is performed.

- To select MAM mode, slide the MAM switch ⑦ upwards to position <<MAM>> until the MAM-symbol ⑳ appears on the display. To change to standard mode (single measurement), slide the MAM switch downwards to position «1».
- The bottom, right hand section of the display shows a 1 or 2 to indicated which of the 2 measurements is currently being taken.
- There is a break of 15 seconds between the measurements. A count down indicates the remaining time.
- The individual results are not displayed. Your blood pressure will only be displayed after all measurements are taken.
- Do not remove the cuff between measurements.
- If one of the individual measurement was questionable, an additional measurement is automatically taken.

### 3. Checklist for Taking a Reliable Measurement

- ▶ Avoid activity, eating or smoking immediately before the measurement.
- ▶ Sit down on a back-supported chair and relax for 5 minutes. Keep the feet flat on the floor and do not cross your legs.
- ▶ **Always measure on the same arm** (normally left). It is recommended that doctors perform double arm measurements on a patients first visit in order to determine which arm to measure in the future. The arm with the higher blood pressure should be measured.
- ▶ Remove close-fitting garments from the upper arm. To avoid constriction, shirt sleeves should not be rolled up - they do not interfere with the cuff if they are laid flat.
- ▶ Always ensure that the correct cuff size is used (marking on the cuff).
  - Fit the cuff closely, but not too tight.
  - Make sure that the cuff is positioned 2 cm above the elbow.
  - The **artery mark** on the cuff (ca. 3 cm long bar) must lie over the artery which runs down the inner side of the arm.
  - Support your arm so it is relaxed.
  - Ensure that the cuff is at the same height as your heart.

### 4. Taking a Blood Pressure Measurement

1. Select standard (single measurement) or MAM mode (automatic double measurement); see details in chapter «2.».
2. Press the ON/OFF button ① to start the measurement.

3. The cuff will now pump up automatically. Relax, do not move and do not tense your arm muscles until the measurement result is displayed. Breathe normally and do not talk.
4. The cuff fit check ㉑ on the display indicates that the cuff is perfectly placed. If the icon ㉑-**A** appears, the cuff is fitted suboptimally, but it is still ok to measure.
5. When the correct pressure is reached, the pumping stops and the pressure falls gradually. If the required pressure was not reached, the device will automatically pump some more air into the cuff.
6. During the measurement, the pulse indicator ㉒ flashes in the display.
7. The result, comprising the systolic ⑩ and the diastolic ⑪ blood pressure and the pulse rate ⑬ is displayed. Note also the explanations on further display symbols in this booklet.
8. When the device has finished measuring, remove the cuff.
9. Switch off the device. (The monitor does switch off automatically after approx. 1 min.).



You can stop the measurement at any time by pressing the ON/OFF button (e.g. if you feel uneasy or an unpleasant pressure sensation).



This monitor is specially tested for use in pregnancy and pre-eclampsia. When you detect unusual high readings in pregnancy, you should measure after a short while again (eg. 1 hour). If the reading is still too high, consult your doctor or gynecologist.  
In pregnancy the AFIB symbol can be ignored.

#### Manual inflation

**In case of high systolic blood pressure (e.g. above 135 mmHg),** it can be an advantage to set the pressure individually. Press the ON/OFF button after the monitor has been pumped up to a level of approx. 30 mmHg (shown on the display). Keep the button pressed until the pressure is about 40 mmHg above the expected systolic value – then release the button.

#### How not to store a reading

As soon as the reading is displayed press and hold the ON/OFF button ① until «**M**» ㉒ is flashing. Confirm to delete the reading by pressing the time button ④.



«**CL**» is displayed when the reading is deleted from the memory successfully.

## How do I evaluate my blood pressure?

The LED traffic light indicator on the left-hand side of the display ⑧ indicates within which range the measured blood pressure lies. The value is either within the optimum (green), elevated (yellow) or high (red) range. The classification corresponds to the following ranges defined by international guidelines (ESH, ESC, JSH). Data in mmHg.

| Range                      | Systolic  | Diastolic | Recommendation      |
|----------------------------|-----------|-----------|---------------------|
| 1. blood pressure too high | ≥135      | ≥85       | Seek medical advice |
| 2. blood pressure elevated | 130 - 134 | 80 - 84   | Self-check          |
| 3. blood pressure optimum  | <130      | <80       | Self-check          |

The higher value is the one that determines the evaluation.

Example: a blood pressure value of **140/80** mmHg or a value of **130/90** mmHg indicates «blood pressure too high».

### Average Indicator «MyCheck»

This symbol ⑩ indicates after each measurement, if the most recent measured value lies below, above or on the same level as your stored average value (see also chapter «5. Data Memory»).

- ➡ If the measured Systole or Diastole is more than 5mmHg higher than the stored average, the arrow shows upwards.
- ➡ If the measured Systole or Diastole is more than 5mmHg lower than the stored average, the arrow shows downwards.
- ➡ If the measured Systole and Diastole do not differ by more than 5mmHg from the stored average, the arrow shows straight on.
- ➡ If the measured systole and diastole differ in different directions from the stored average, this is indicated first with the systole figure flashing, together with the up or down arrow for two seconds. Thereafter, the diastole figure flashes with the arrow pointing up or down for two seconds.

### Appearance of the Pulse Arrhythmia (PAD)

The pulse arrhythmia indicator ⑪ is displayed, when pulse irregularities occurred during the measurement. If the symbol appears, it is recommended to select MAM mode and repeat measurement to double confirm the blood pressure reading: see details in chapter «<<2>>».

- ➡ When Atrial Fibrillation (AF) is detected, the PAD is also indicated because AF is one type of pulse irregularities.

## 5. Data Memory

This device automatically stores up to 99 measurement values for each of the 2 users.

Select either user 1 or 2 by pressing the user button ⑤.

### Viewing the average of the last 28 days

Press the M-button ③ briefly, when the device is switched off. The display first shows «M» ②⑥ and «28A», which stands for the average of measurement values of the last 28 days.

- ➡ Blood pressure readings with suboptimal cuff fit ②⑦-A are not considered in the average value.

### Viewing the clinical blood pressure average «MyBP»

Pressing the M-button again, allows you to see the clinically relevant blood pressure average «MyBP». The display first shows «M» ②⑥ and «MyBP» ②⑦. This clinical blood pressure average is only shown when 12 clinically relevant measurement values in the last 28 days have been performed.

- ➡ Only measurements that were performed in the morning between 05:00-10:59 or in the evening between 17:00-22:59 are considered.
- ➡ A maximum of 4 readings per day are considered (2 from the morning and 2 from the evening).
- ➡ Measurements performed in standard mode and in MAM-mode are considered in the average, when they have been performed during the right time of the day.
- ➡ Measurements performed in MAM-mode or single standard mode are both classed as single measurements for working out the «MyBP average».
- ➡ Blood pressure readings with suboptimal cuff fit ②⑦-A are not considered in the average value.

### Viewing the stored single values

Pressing the M-button again, allows you to see the last performed measurement. The display first shows «M» ②⑥ and a value, e.g. «M17». This means that there are 17 single values in the memory. Pressing the M-button again displays the previous value. Pressing the M-button repeatedly enables you to move from one stored value to another.

- ➡ Pay attention that the maximum memory capacity of 99 memories is not exceeded. **When the 99 memory is full, the oldest value is automatically overwritten with the 100th**

**value.** Values should be evaluated by a doctor before the memory capacity is reached – otherwise data will be lost.

### Clearing all values

Make sure the correct user is activated.

If you are sure that you want to permanently remove all stored values, hold down the M-button (the device must have been switched off beforehand) until «CL ALL» appears and then release the button. To permanently clear the memory, press the time button while «CL ALL» is flashing. Individual values cannot be cleared.

 **Cancel deletion:** press ON/OFF button ① while «CL ALL» is flashing.

## 6. Battery Indicator and Battery change

---

### Low battery

When the batteries are approximately ¼ empty the battery symbol ⑱ will flash as soon as the device is switched on (partly filled battery displayed). Although the device will continue to measure reliably, you should obtain replacement batteries.

### Flat battery – replacement

When the batteries are flat, the battery symbol ⑲ will flash as soon as the device is switched on (flat battery displayed). You cannot take any further measurements and must replace the batteries.

1. Open the battery compartment ⑩ at the back of the device.
2. Replace the batteries – ensure correct polarity as shown by the symbols in the compartment.
3. To set date and time, follow the procedure described in «Section 2.».

 The memory retains all values although date and time must be reset – the year number therefore flashes automatically after the batteries are replaced.

### Which batteries and which procedure?

-  Use 4 new, long-life 1.5 V, size AA alkaline batteries.
-  Do not use batteries beyond their date of expiry.
-  Remove batteries if the device is not going to be used for a prolonged period.

### Using rechargeable batteries

You can also operate this device using rechargeable batteries.

 Only use «NiMH» type reusable batteries.

 Batteries must be removed and recharged when the flat battery symbol appears. They should not remain inside the device as they may become damaged (total discharge as a result of low use of the device, even when switched off).

 Always remove the rechargeable batteries if you do not intend to use the device for a week or more.

 Batteries cannot be charged in the blood pressure monitor. Recharge batteries in an external charger and observe the information regarding charging, care and durability.

## 7. Using a Mains Adapter

---

You can operate this device using the Microlife mains adapter (DC 6V, 600 mA).

 Only use the Microlife mains adapter available as an original accessory appropriate for your supply voltage.

 Ensure that neither the mains adapter nor the cable are damaged.

1. Plug the adapter cable into the mains adapter socket ⑩ in the blood pressure monitor.
2. Plug the adapter plug into the wall socket.

When the mains adapter is connected, no battery current is consumed.

## 8. Bluetooth® Function

---

This device can be used in conjunction with a smartphone running the «Microlife Connected Health+» App. The Bluetooth® connection is automatically active ⑮ after the measurement has been completed.

 To manually activate the Bluetooth®, press ON/OFF ① for 4 seconds until the Bluetooth symbol starts flashing ⑮.

For more detailed information visit [www.microlife.com/connect](http://www.microlife.com/connect).

## 9. PC-Link Functions

---

This device can be used in conjunction with a personal computer (PC) running the Microlife Blood Pressure Analyzer+ (BPA+) software. The memory data can be transferred to the PC by connecting the monitor via a cable.

If no download-voucher and cable is included download the BPA software from [www.microlife.com/technologies/connect/bpaplus](http://www.microlife.com/technologies/connect/bpaplus) and use a Micro-USB cable.

## 10. Error Messages

If an error occurs during the measurement, the measurement is interrupted and an error message, e.g. «Err 3», is displayed.

| Error  | Description                        | Potential cause and remedy  |
|--|------------------------------------|---|
| «Err 1»<br>⑰   | Signal too weak                    | The pulse signals on the cuff are too weak. Re-position the cuff and repeat the measurement.*   |
| «Err 2»<br>⑰-B   | Error signal                       | During the measurement, error signals were detected by the cuff, caused for instance by movement or muscle tension. Repeat the measurement, keeping your arm still.   |
| «Err 3»<br>⑰-C   | Abnormal cuff pressure             | An adequate pressure cannot be generated in the cuff. A leak may have occurred. Check that the cuff is correctly connected and is not too loose. Replace the batteries if necessary. Repeat the measurement.                          |
| «Err 5»  | Abnormal result                    | The measuring signals are inaccurate and no result can therefore be displayed. Read through the checklist for performing reliable measurements and then repeat the measurement.*  |
| «Err 6»  | MAM Mode                           | There were too many errors during the measurement in MAM mode, making it impossible to obtain a final result. Read through the checklist for performing reliable measurements and then repeat the measurement.*                       |
| «HI»   | Pulse or cuff pressure too high    | The pressure in the cuff is too high (over 299 mmHg) OR the pulse is too high (over 200 beats per minute). Relax for 5 minutes and repeat the measurement.*   |
| «LO»   | Pulse too low                      | The pulse is too low (less than 40 beats per minute). Repeat the measurement.*  |
|  | Problem with Bluetooth® connection | If any problem occurs with the Bluetooth connection, the Bluetooth® icon ⑱ blinks rapidly for approximately 10 seconds. To solve the problem, please visit <a href="http://www.microlife.com/connect">www.microlife.com/connect</a> . |

\* Please immediately consult your doctor, if this or any other problem occurs repeatedly.

## 11. Safety, Care, Accuracy Test and Disposal



### Safety and protection

- Follow instructions for use. This document provides important product operation and safety information regarding this device. Please read this document thoroughly before using the device and keep for future reference.
- This device may only be used for the purposes described in these instructions. The manufacturer cannot be held liable for damage caused by incorrect application.
- This device comprises sensitive components and must be treated with caution. Observe the storage and operating conditions described in the «Technical Specifications» section.
- Protect it from:
  - water and moisture
  - extreme temperatures
  - impact and dropping
  - contamination and dust
  - direct sunlight
  - heat and cold
- The cuffs are sensitive and must be handled with care.
- Do not exchange or use any other kind of cuff or cuff connector for measuring with this device.
- Only pump up the cuff once fitted.
- Do not use this device close to strong electromagnetic fields such as mobile telephones or radio installations. Keep a minimum distance of 3.3 m from such devices when using this device.
- Do not use this device if you think it is damaged or notice anything unusual.
- Never open this device.
- If the device is not going to be used for a prolonged period the batteries should be removed.
- Read the additional safety information provided within the individual sections of this instruction manual.
- The measurement results given by this device is not a diagnosis. It is not replacing the need for the consultation of a physician, especially if not matching the patient's symptoms. Do not rely on the measurement result only, always consider other potentially occurring symptoms and the patient's feedback. Calling a doctor or an ambulance is advised if needed.

- **Permanently high blood pressure values can damage your health and must be treated by your doctor!**
- Always discuss your values with your doctor and tell him/her if you have noticed anything unusual or feel unsure. **Never rely on single blood pressure readings.**
- **Under no circumstances should you alter the dosages of drugs or initiate a treatment without consulting your doctor.**
- **Deviations** between measurements taken by your doctor or in the pharmacy and those taken at home are quite normal, as these situations are completely different.
- **The pulse display is not suitable for checking the frequency of heart pacemakers!**
- If you are **pregnant**, you should monitor your blood pressure regularly as it can change drastically during this time.



Ensure that children do not use this device unsupervised; some parts are small enough to be swallowed. Be aware of the risk of strangulation in case this device is supplied with cables or tubes.

#### Device care

Clean the device only with a soft, dry cloth.

#### Cleaning the cuff

Carefully remove spots on the cuff with a damp cloth and soapsuds.



**WARNING:** Under no circumstances may you wash the inner bladder!

#### Accuracy test

We recommend this device is tested for accuracy every 2 years or after mechanical impact (e.g. being dropped). Please contact your local Microlife-Service to arrange the test (see foreword).

#### Disposal



Batteries and electronic devices must be disposed of in accordance with the locally applicable regulations, not with domestic waste.

## 12. Guarantee

This device is covered by a **5 year guarantee** from the date of purchase. The guarantee is valid only on presentation of the guarantee card completed by the dealer (see back) confirming date of purchase or the receipt.

- Batteries and parts that become worn with use are not included.
- Opening or altering the device invalidates the guarantee.

- The guarantee does not cover damage caused by improper handling, discharged batteries, accidents or non-compliance with the operating instructions.
- The cuff has a functional guarantee (bladder tightness) for 2 years. Please contact your local Microlife-Service (see foreword).

## 13. Technical Specifications

**Operating conditions:** 10 - 40 °C / 50 - 104 °F  
15 - 95 % relative maximum humidity

**Storage conditions:** -20 - +55 °C / -4 - +131 °F  
15 - 95 % relative maximum humidity

**Weight:** 415 g (including batteries)

**Dimensions:** 157.5 x 105 x 61.5 mm

**Measuring procedure:** oscillometric, corresponding to Korotkoff method: Phase I systolic, Phase V diastolic

**Measurement range:** 20 - 280 mmHg – blood pressure  
30 - 199 beats per minute – pulse

**Cuff pressure display range:** 0 - 299 mmHg

**Resolution:** 1 mmHg

**Static accuracy:** pressure within  $\pm 3$  mmHg

**Pulse accuracy:**  $\pm 5$  % of the readout value

**Voltage source:** 4 x 1.5 V alkaline batteries; size AA  
Mains adapter DC 6V, 600 mA (optional)

**Battery lifetime:** approx. 920 measurements  
(using new batteries)

**IP20:** Protected against solid foreign particles with a diameter of more than 12.5 mm, no protection against water

**Reference to standards:** ANSI/AAMI/IEC 80601-2-30;  
ANSI/AAMI/ISO 81060-2; IEC 60601-1;  
IEC 60601-1-2 (EMC); IEC 60601-1-11

**Expected service life:** Device: 5 years or 10000 measurements  
Cuff: 2 years or 10000 measurements

This device complies with the requirements of the Medical Device Directive 93/42/EEC.  
Technical alterations reserved.

The Bluetooth® word mark and logos are registered trademarks owned by Bluetooth SIG, Inc. and any use of such marks by Micro-life Corp. is under license. Other trademarks and trade names are those of their respective owners.